

Schedule

Issue date: 18 May 2018
Valid until: 23 November 2020



NO: SAMM 082

(Issue 2, 18 May 2018 replacement of SAMM 082 dated 8 December 2017)

Page: 1 of 40

LABORATORY LOCATION: (PERMANENT LABORATORY)



SENDI MAHIR SDN. BHD.
NO. 6, 8 & 10,12 JALAN KAPAR 27/89
MEGAH INDUSTRIAL PARK
40400 SHAH ALAM, SELANGOR
MALAYSIA

FIELDS OF CALIBRATION:

FORCE, TORQUE, DIMENSIONAL, MASS, FLOW, PRESSURE, TEMPERATURE, VOLUMETRIC, ELECTRICAL & OPTICAL AND PHOTOMETRIC MEASUREMENTS

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2005 (ISO/IEC 17025:2005).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

* The expanded uncertainties are based on an estimated confidence probability of approximately 95% and have a coverage factor of $k=2$ unless stated otherwise.

SCOPE OF CALIBRATION: FORCE

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Push-Pull Gauge	0 kgf to 100 kgf	0.5 % of reading	Calibrated using Deadweight Standard Weight based on ISO 376:2004
Tension Gauge	0 kgf to 100 kgf	0.5 % of reading	
Load Measuring Device			Calibrated using Load Cell, Proving Ring and Tension/ Compression Testing Machine based on ISO 376:2004
Tension	0 kgf to 500 kgf	15 gf	
	500 kgf to 1 tonf	0.58 kgf	
	1 tonf to 5 tonf	1.7 kgf	
	5 tonf to 10 tonf	13 kgf	
	10 tonf to 30 tonf	79 kgf	
Compression	0 kgf to 500 kgf	15 gf	
	500 kgf to 1 tonf	0.56 kgf	
	1 tonf to 5 tonf	1.7 kgf	
	5 tonf to 10 tonf	13 kgf	
	10 tonf to 30 tonf	79 kgf	
	30 tonf to 40 tonf	93 kgf	

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)**SCOPE OF CALIBRATION: FORCE**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Shore Hardness Tester (spring load)			
1. Type A, B, E, O	0 to 100 shore hardness index	0.24 shore hardness index	Calibrated using Durometer tester based on ASTM D 2240:2005
2. Type C, D, DO	0 to 100 shore hardness index	0.23 shore hardness index	
3. Type OO	0 to 100 shore hardness index	0.23 shore hardness index	
Adhesion Tester (pressure)	0 N/mm ² to 25 N/mm ²	0.14 N/mm ²	Calibrated using Load Cell based on ASTM D 4541:2009

Signatory:

1. **Seah Leong Ho**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)**SCOPE OF CALIBRATION: FORCE****SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Universal Testing Machine 1. Tensile mode (0~10,000 kgf) 2. Compress mode (0~200,000 kgf)	0 kgf to 500 kgf 500 kgf to 1,000 kgf 1,000 kgf to 5,000 kgf 5,000 kgf to 10,000 kgf 10,000 kgf to 50,000 kgf 50,000 kgf to 200,000 kgf	24 gf 0.26 kgf 1 kgf 15 kgf 60 kgf 260 kgf	Calibrated using Deadweight up to 500 kgf or Load Cell based on ISO 7500-1:2004
Hardness Tester	20 HRA to 88 HRA 30 HRB to 100 HRB 10 HRC to 70 HRC	\pm 0.6 HRA \pm 0.6 HRB \pm 0.6 HRC	Calibrated using Load Cell and Hardness Block set based on ISO 6508-2:2005 under method clause 5 indirect verification

Signatory:

1. **Seah Leong Ho**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)**SCOPE OF CALIBRATION: TORQUE**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Torque Measuring Device	0 N.m to 1 N.m 1 N.m to 10 N.m 10 N.m to 100 N.m 100 N.m to 500 N.m 500 N.m to 1,000 N.m	0.52 % of reading 0.33 % of reading 0.19 % of reading 0.15 % of reading 0.15 % of reading	Calibrated using ERM and Torque Transducer or Torque Bar based on BS 7882:2008
Torque Tools Device (Torque wrench and Torque Driver)	0. N.m to 1 N.m 1 N.m to 10 N.m 10 N.m to 100 N.m 100 N.m to 500 N.m 500 N.m to 1,000 N.m 1,000 N.m to 3,000 N.m	0.52 % of reading 0.33 % of reading 0.19 % of reading 0.15 % of reading 0.15 % of reading 0.26 % of reading	Calibrated using ERM and Torque Transducer or Torque Bar based on ISO 6789-1:2017 (E)

Signatory:

1. **Seah Leong Ho**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)**SCOPE OF CALIBRATION: DIMENSIONAL**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Plain Plug Gauge / Pin Gauge (Diameter)	0 mm to 10 mm	0.2 μ m	Calibrated using ULM
	10 mm to 100 mm	0.4 μ m	
	100 mm to 300 mm	1.0 μ m	
Plain Ring Gauge (Diameter)	0 mm to 10 mm	0.4 μ m	Calibrated using ULM
	10 mm to 100 mm	0.5 μ m	
	100 mm to 300 mm	1.0 μ m	
External Micrometer 50 mm frame 100 mm frame 150 mm frame 200 mm frame 300 mm frame 400 mm frame 450 mm frame 600 mm frame	Up to 25 mm	1 μ m	Calibrated using Gauge Blocks based on BS EN ISO 3611:2010
	25 mm tranverse	1.1 μ m	
	25 mm tranverse	1.2 μ m	
	25 mm tranverse	1.3 μ m	
	25 mm tranverse	1.6 μ m	
	25 mm tranverse	2 μ m	
	25 mm tranverse	2.5 μ m	
25 mm tranverse	2.5 μ m		
Caliper Checker	0 mm to 600 mm	3 μ m	Calibrated using Gauge Blocks based on ISO 7863:1984
Dial / Digimatic & Vernier Caliper	0 mm to 200 mm	5 μ m	Calibrated using Gauge Blocks based on BS EN ISO 13385- 1:2011
	200 mm to 450 mm	7 μ m	
	450 mm to 1,000 mm	7.5 μ m	
	1,000 mm to 1,500 mm	17 μ m	
	1,500 mm to 2,000 mm	23 μ m	
Dial Gauge	0 mm to 50 mm	1.5 μ m	Calibrated using Dial Gauge Calibrator based on BS 907:2008
Dial Test Indicator	0 mm to 50 mm	1.5 μ m	Calibrated using Dial Gauge Calibrator based on BS 2795:1981
Dial / Digimatic & Vernier Height Gauge	0 mm to 450 mm	7 μ m	Calibrated using Gauge Blocks based on BS EN ISO 13225:2012
	450 mm to 600 mm	7.5 μ m	
	600 mm to 1,000 mm	7.7 μ m	
Height Setting Micrometer & Riser Block	0 mm to 600 mm	3 μ m	Calibrated using Gauge Blocks based on ISO 7863:1984

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)**SCOPE OF CALIBRATION: DIMENSIONAL**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Screw Plug (simple pitch diameter)	0 mm to 20 mm	0.5 μ m	Calibrated using ULM
	20 mm to 100 mm	1.0 μ m	
Screw Ring (simple pitch diameter)	0 mm to 20 mm	0.5 μ m	Calibrated using ULM
	20 mm to 100 mm	1.0 μ m	
Gauge Block Set Grade '0' and below	0 mm to 10 mm	0.12 μ m	Calibrated using Gauge Blocks based on ISO 3650:1998
	10 mm to 25 mm	0.12 μ m	
	25 mm to 50 mm	0.13 μ m	
	50 mm to 75 mm	0.14 μ m	
	75 mm to 100 mm	0.16 μ m	
Steel Rulers	0 mm to 1,000 mm	0.2 mm	Calibrated using Standard Scale based on JIS B 7516:2005
	1,000 mm to 3,000 mm	0.5 mm	
Vee Blocks	220 mm x 160 mm x 80 mm	4 μ m	Calibrated using Dial Test Indicator based on JIS B 7540:1972
Bubble Levelling Gauge	Height in respect to length 0.02 mm/m to 0.25 mm/m	6 μ m/m	Calibrated using Bubble Tube Tester and Dial Test Indicator based on JIS B 7510:1993
Cylinder Gauge	0 mm to 600 mm	1 μ m	Calibrated using ULM and Gauge Blocks based on JIS B 7515:1982
Internal Micrometer	0 mm to 25 mm	1.5 μ m	Calibrated using ULM and Gauge Blocks based on BS 959:2008
	25 mm to 100 mm	2 μ m	
	100 mm to 600 mm	10 μ m	
Setting Rod	0 mm to 25 mm	0.3 μ m	Calibrated using ULM and Gauge Blocks based on BS 870:2008
	25 mm to 100 mm	0.5 μ m	
	100 mm to 600 mm	2.2 μ m	

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)**SCOPE OF CALIBRATION: DIMENSIONAL**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Thickness Coating Film	0 mm to 2.5 mm	0.2 μ m	Calibrated using ULM and Gauge Blocks based on BS 5411
Pitch Gauge	0 mm to 12 mm	10 μ m	Calibrated using Profile Projector
Radius Gauge	0 mm to 100 mm	10 μ m	Calibrated using Profile Projector
Test Sieves	0 mm to 200 mm	10 μ m	Calibrated using Profile Projector
M μ Checker	0 mm to 3 mm	0.2 μ m	Calibrated using Gauge Blocks based on JIS B 7536 : 1982
Feeler Gauge (Thickness)	0.01 mm to 10 mm	2 μ m	Calibrated using Precision Micrometer based on BS 957:2008
Dial Gauge Calibrator	0 mm to 25 mm	0.2 μ m	Calibrated using Precision Digital Linear Probe based on JIS B 7519:1994
Depth Gauge	0 mm to 25 mm 25 mm to 100 mm 100 mm to 300 mm	1.4 μ m 1.5 μ m 2.0 μ m	Calibrated using Gauge Blocks based on BS 6468:2008
Thickness Gauge	0 mm to 10 mm 10 mm to 65 mm	1.2 μ m 1.4 μ m	Calibrated using Gauge Blocks based on JIS B 7519:1994
Snap Gauge	0 mm to 100 mm 100 mm to 200 mm	0.4 μ m 0.6 μ m	Calibrated using ULM and Setting Ring

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)**SCOPE OF CALIBRATION: DIMENSIONAL**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
<u>Taper Plain Plug</u> Diameter	0 mm to 20 mm 20 mm to 300 mm	0.5 μ m 1.0 μ m	Calibrated using ULM and Setting Ring
Angle		0.003°	Calibrated using ULM
<u>Taper Plain Ring</u> Diameter	0 mm to 20 mm 20 mm to 300 mm	0.5 μ m 1.0 μ m	Calibrated using ULM and Setting Ring
Angle		0.003°	Calibrated using ULM
<u>Taper Thread Plug</u> Pitch Diameter	0 mm to 20 mm 20 mm to 100 mm	0.5 μ m 1.0 μ m	Calibrated using ULM
Angle		0.002°	Calibrated using ULM
<u>Taper Thread Ring</u> Pitch Diameter	0 mm to 20 mm 20 mm to 100 mm	0.5 μ m 1.0 μ m	Calibrated using ULM and Setting Ring
Angle		0.002°	Calibrated using ULM
Length Bar	125 mm 150 mm 175 mm 200 mm 250 mm 300 mm 400 mm 500 mm	0.63 μ m 0.71 μ m 0.79 μ m 0.87 μ m 1.0 μ m 1.2 μ m 1.6 μ m 2.3 μ m	Calibrated using ULM and Gauge Blocks

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)**SCOPE OF CALIBRATION: DIMENSIONAL**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Tape	1. Steel type Up to 1,000 mm 2,000 mm 5,000 mm 8,000 mm 10,000 mm 20,000 mm 30,000 mm 50,000 mm 100,000 mm	0.12 mm	Calibrated using scale and tape calibration unit based on JIS B 7512:1993
		0.17 mm	
0.26 mm			
0.33 mm			
0.37 mm			
0.53 mm			
0.64 mm			
0.83 mm			
1.2 mm			
2. Fabric type	Up to 1,000 mm 2,000 mm 5,000 mm 8,000 mm 10,000 mm 20,000 mm 50,000 mm 100,000 mm	0.12 mm	Calibrated using scale and tape calibration unit based on JIS B 7522:2005
		0.17 mm	
		0.27 mm	
		0.34 mm	
		0.38 mm	
		0.53 mm	
		0.84 mm	
		1.2 mm	
Holtest	0 mm to 175 mm 175 mm to 200 mm	1 μ m	Calibrated using Master Ring Gauge based on DIN 863-4:1999 for repeatability test only
		2 μ m	
Bevel Protractor	0° to 360°	0.6°	Calibrated using inclinometer and Feeler Gauge based on BS 1685:2008

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation**Signatories:**

1. **Seah Leong Ho**
2. **Chin Inn Nkot**
3. **Kayalvili a/p Munusamy**
4. **Norita binti Md. Ali**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 10 of 40

SCOPE OF CALIBRATION: DIMENSIONAL**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Projector (individual linear axis only)	0 mm to 50 mm	1.7 μ m Magnification 0.1%	Calibrated using Glass Scale, Precision Ball and Reading Scale based on JIS B7184:1999
	50 mm to 100 mm	1.9 μ m Magnification 0.1%	
	100 mm to 200 mm	2.5 μ m Magnification 0.1%	
	200 mm to 300 mm	3.5 μ m Magnification 0.1%	
Co-Ordinate Measuring Machine	0 mm to 1,000 mm	10 μ m	Calibrated using Ball Bar Set, Long Gauge Block, Gauge Block Set and Thermometer with Sensor based on ANSI/ASME B89:1997
Caliper	0 mm to 1,000 mm	7.5 μ m	Calibrated using Gauge Blocks based on BS EN ISO 13385- 1:2011
	1,000 mm to 1,500 mm	17 μ m	
	1,500 mm to 2,000 mm	23 μ m	
Micrometer (External)	0 mm to 600 mm	2.5 μ m	Calibrated using Gauge Blocks based on BS EN ISO 3611-2010
Linear Height Gauge	0 mm to 1,000 mm	7.5 μ m	Calibrated using Gauge Blocks based on BS EN ISO 13225 : 2012

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 11 of 40

SCOPE OF CALIBRATION: DIMENSIONAL**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Surface Plate Flatness	600 mm x 600 mm 800 mm x 800 mm 1m x1m	1.3 μ m 1.8 μ m 2.2 μ m	Calibrated using Planekator, Repeat-O-Meter and Micro- Comparator based on BS 817:2008

Signatories:

1. Seah Leong Ho
2. Chin Inn Nkot

SCOPE OF CALIBRATION: DIMENSIONAL**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Microscope (Individual linear axis only)	0 mm to 200 mm	2.2 μ m	Calibrated using Glass Scale based on JIS B 7153:1995

Signatories:

1. Seah Leong Ho

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 12 of 40

SCOPE OF CALIBRATION: MASS

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Standard Weight	1 mg	3 μ g	Calibrated using Standard Weight Sets and Mass Comparator based on OIML R111-2:2004
	2 mg	3 μ g	
	5 mg	3 μ g	
	10 mg	3 μ g	
	20 mg	3 μ g	
	50 mg	3 μ g	
	100 mg	3 μ g	
	200 mg	3 μ g	
	500 mg	4 μ g	
	1 g	4 μ g	
	2 g	7 μ g	
	5 g	10 μ g	
	10 g	20 μ g	
	20 g	30 μ g	
	50 g	60 μ g	
	100 g	0.13 mg	
	200 g	0.26 mg	
	500 g	0.67 mg	
	1 kg	1.3 mg	
	2 kg	2.6 mg	
	5 kg	6.7 mg	
	10 kg	26 mg	
	20 kg	52 mg	
	25 kg	0.15 g	

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation**Signatory:**1. **Seah Leong Ho**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 13 of 40

SCOPE OF CALIBRATION: MASS**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Analytical Balance and Weighing Scale	0 g to 5 g	20 μ g	Calibrated using Standard Weight Sets based on ASTM E898:2005
	0 g to 20 g	20 μ g	
	0 g to 50 g	0.15 mg	
	0 g to 100 g	0.11 mg	
	0 g to 200 g	0.20 mg	
	0 g to 500 g	0.50 mg	
	0 kg to 1 kg	1.0 mg	
	0 kg to 2 kg	2.0 mg	
	0 kg to 3 kg	2.2 mg	
	0 kg to 5 kg	5.0 mg	
	0 kg to 10 kg	47 mg	
	0 kg to 20 kg	96 mg	
	0 kg to 30 kg	1.1 g	
	0 kg to 50 kg	2.0 g	
	0 kg to 100 kg	5.0 g	
	0 kg to 200 kg	9.0 g	
	0 kg to 500 kg	20 g	
	0 kg to 750 kg	45 g	
	0 kg to 1,000 kg	47 g	
	0 kg to 1,500 kg	54 g	
0 kg to 2,000 kg	94 g		
Standard Weight	1 kg	14 mg	Calibrated using Standard Weight Sets and Mass Comparator based on OIML R111-2:2004
	2 kg	14 mg	
	5 kg	16 mg	
	10 kg	0.14 g	
	20 kg	0.15 g	
	25 kg	0.15 g	

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation**Signatory:**

- Seah Leong Ho**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 14 of 40

SCOPE OF CALIBRATION: PRESSURE

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
<u>Pressure Measuring Device</u>			
Vacuum	0 bar to -0.95 bar	0.42 mbar	Calibrated using Multifunction Calibrator based on AS 1349:1987
Pneumatic	0 bar to 20 bar	7 mbar	
Hydraulic	1 bar to 700 bar	0.015% of reading	Calibrated using Deadweight Tester based on AS 1349:1987
	700 bar to 1,100 bar	0.015% of reading	

Signatories:

1. Seah Leong Ho
2. Shahizan Hardy bin Abd Razak

SCOPE OF CALIBRATION: PRESSURE**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Pressure Measuring Device	-700 mbar to 700 mbar 0 bar to 20 bar 0 psi to 11,000 psi	1.7 mbar 50 mbar 0.025 % of reading	Calibrated using Pressure Calibrator based on AS 1349:1987

Signatory:

1. Seah Leong Ho

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 15 of 40

SCOPE OF CALIBRATION: FLOW

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Anemometer	0.15 m/s to 5 m/s	0.12 m/s	Calibrated using Anemometer Calibrator
	5 m/s to 20 m/s	0.24 m/s	
	20 m/s to 45 m/s	0.54 m/s	

Signatory:

1. **Seah Leong Ho**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 16 of 40

SCOPE OF CALIBRATION: TEMPERATURE

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Temperature Measuring Device	-20 °C to 50 °C	0.2° C	Calibrated using Temperature Recorder, PRT, Humidity Chamber and Thermohygrometer based on JIS B7306:1989 and JIS Z8806:1995
Relative Humidity Measuring Device	20 %RH to 95 %RH	2 %RH	
Liquid-In-Glass Thermometer (Total & Partial Immersion)	-30 °C to 0 °C 0 °C to 250 °C 250 °C to 400 °C	0.024 °C 0.041 °C 0.11 °C	Calibrated using Resistance Thermometer Display & PRT, Temperature Recorder, Constant Stirred Low Temperature Liquid Bath, Constant Stirred High Temperature Liquid Bath, Metal Block Bath and Ice Point based on ASTM E77:2007
Temperature Sensor (Thermocouple, PRT, Mechanical Thermometer & Thermistor)	-95 °C to -50 °C -50 °C to -30°C -30 °C to 0 °C 0 °C to 250 °C 250 °C to 400 °C 400 °C to 660 °C 660 °C to 800 °C 800 °C to 1,000 °C 1,000 °C to 1,200 °C	0.063 °C 0.062 °C 0.024 °C 0.041 °C 0.11 °C 0.13 °C 2.1 °C 3.1 °C 4.0 °C	Calibrated using Resistance Thermometer Display & PRT, Type R Thermocouple, Temperature Recorder, Constant Stirred Low Temperature Liquid Bath, Constant Stirred High Temperature Liquid Bath, Temperature Block Calibrator and Ice Point based on JIS C1602:1995, JIS C 1604:1997, JIS C1611:1995 and JIS Z 8710:1993

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)**SCOPE OF CALIBRATION: TEMPERATURE**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Temperature Block Calibrator	-95 °C to -50 °C -50 °C to 660 °C 660 °C to 1,200 °C	0.057 °C 0.013 °C 2.7 °C	Calibrated using PRT, Thermocouple Type R and Temperature Display based on EA-10/13:1999
Liquid Bath	-50 °C to 300 °C	0.012 °C	
Temperature Indicating Device Resistance Type PT100	-200 °C to 650 °C 650 °C to 850 °C	0.1 °C 0.1 °C	Calibrated by electrical simulation using Calibrator and Ice point based on JIS C 1601:1983 and JIS C 1603:1983
Thermocouple Type			
K	-270 °C to -100 °C -100 °C to 1,370 °C	1.5 °C 0.1 °C	
J	-210 °C to 1,200 °C	0.1 °C	
T	-270 °C to -100 °C -100 °C to 400 °C	0.5 °C 0.1 °C	
E	-270 °C to -100 °C -100 °C to 1,000 °C	0.3 °C 0.1 °C	
R	0 °C to 500 °C 500 °C to 1,760 °C	0.4 °C 0.1 °C	
S	0 °C to 500 °C 500 °C to 1,760 °C	0.4 °C 0.1 °C	
N	-100 °C to 1,300 °C	0.1 °C	

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 18 of 40

SCOPE OF CALIBRATION: TEMPERATURE

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Temperature Calibrator Resistance Type PT100	-200 °C to 650 °C	0.1 °C	Calibrated by electrical measurement using Multimeter and Ice Point based on JIS Z 8704:1993 and ITS 90:1990
JPT100	650 °C to 850 °C	0.1 °C	
	-200 °C to 500 °C	0.1 °C	
Thermocouple Type			
K	-270 °C to -100 °C	1.5 °C	
	-100 °C to 1,370 °C	0.1 °C	
J	-210 °C to 1,200 °C	0.1 °C	
T	-270 °C to -100 °C	0.5 °C	
	-100 °C to 400 °C	0.1 °C	
E	-270 °C to -100 °C	0.3 °C	
	-100 °C to 1,000 °C	0.1 °C	
N	-270 °C to -100 °C	1.5 °C	
	-100 °C to 1,300 °C	0.1 °C	
B	100 °C to 500 °C	1.5 °C	
	500 °C to 1,820 °C	0.3 °C	
R	0 °C to 500 °C	0.4 °C	
	500 °C to 1,760 °C	0.2 °C	
S	0 °C to 500 °C	0.4 °C	
	500 °C to 1,760 °C	0.2 °C	
Radiation Thermometer	-20 °C to 150 °C	0.39 °C	Calibrated using Temperature Recorder, Thermocouple, Single Blackbody Calibrator and Cyclops Spherical Blackbody Source based on ASTM E 1256:2007
	150 °C to 200 °C	0.93 °C	
	200 °C to 400 °C	0.99 °C	
	400 °C to 1,000 °C	2.0 °C	
	1,000 °C to 1,300 °C	3.6 °C	

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation**Signatories:**

1. **Seah Leong Ho**
2. **Mohamad Azlan bin Mohamed Aris**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 19 of 40

SCOPE OF CALIBRATION: TEMPERATURE**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Temperature Sensor (Thermocouple, PRT, Mechanical Thermometer & Thermistor)	-30 °C to 30 °C 30 °C to 650 °C	0.051 °C 0.17 °C	Calibrated using Resistance Thermometer Display & PRT, Temperature Recorder, Constant Stirred Temperature Liquid Bath, Constant Stirred High Temperature Liquid Bath, Temperature Block Calibrator and Ice Point based on JIS C 1602:1995, JIS C 1604:1997, JIS C 1611:1995 and JIS Z 8710:1993
Temperature Controlled Enclosure	-80 °C to -20 °C -20 °C to 250 °C 250 °C to 800 °C 800 °C to 1,000 °C 1,000 °C to 1,200 °C	0.6 °C 1 °C 2 °C 3 °C 5 °C	Calibrated using Temperature Recorder, PRT and Thermocouple based on AS 2853:1986
Liquid Bath	-30 °C to 300 °C	0.1 °C	
Humidity Controlled Enclosure	30 %RH to 98 %RH	5 %RH	Calibrated using Temperature Recorder, PRT, Thermocouple and Barometer based on AS 2853:1986 and BS 1339-3:2004

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 20 of 40

SCOPE OF CALIBRATION: TEMPERATURE**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Temperature Indicating Device Resistance Type PT100 JPT100	-200 °C to 650 °C -200 °C to 500 °C	0.1 °C 0.1 °C	Calibrated by electrical simulation using Calibrator and Ice Point based on JIS C1601:1983 and JIS C1603:1983
Thermocouple Type K	-270 °C to -100 °C -100 °C to 1,370 °C	1.5 °C 1 °C	
J	-210 °C to 1,200 °C	1 °C	
T	-270 °C to 400 °C	1 °C	
E	-270 °C to 1,000 °C	1 °C	
R	0 °C to 1,760 °C	1 °C	
S	0 °C to 1,760 °C	1 °C	
N	-100 °C to 1,300 °C	1 °C	

Signatories:

1. **Seah Leong Ho**
2. **Mohd Safiee bin Ngadirin**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 21 of 40

SCOPE OF CALIBRATION: VOLUMETRIC

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Burette	1 ml to 10 ml 10 ml to 25 ml 25 ml to 50 ml 50 ml to 100 ml	6 μ l 20 μ l 30 μ l 70 μ l	Calibrated using Analytical Balance and Distilled Water based on ISO 385:2005(E)
Measuring Cylinder	5 ml 5 ml to 10 ml 10 ml to 25 ml 25 ml to 100 ml 100 ml to 250 ml 250 ml 500 ml 500 ml to 1,000 ml 1,000 ml to 2,000 ml	40 μ l 70 μ l 0.2 ml 0.4 ml 0.7 ml 1.6 ml 3 ml 5 ml	Calibrated using Analytical Balance and Distilled Water based on ISO 4788:2005(E)
One Mark Volumetric Flask	5 ml to 10 ml 10 ml to 25 ml 25 ml to 100 ml 100 ml to 500 ml 500 ml 2,000 ml	20 μ l 30 μ l 60 μ l 0.2 ml 0.3 ml	Calibrated using analytical Balance and Distilled Water based on ISO 1042:1998
Pipette	Type 1, Type 2, Type 3 0.5 ml to 1 ml 1 ml to 2 ml 2 ml to 5 ml 5 ml to 10 ml 10 ml to 25 ml 25 ml to 100 ml	4 μ l 7 μ l 15 μ l 29 μ l 57 μ l 7 μ l	Calibrated using Analytical Balance and Distilled Water based on ISO 835:2007(E) Calibrated using Analytical Balance and Distilled Water based on ISO 4787:2010
Piston Operated Volumetric Apparatus (POVA)	10 μ l to 200 μ l 200 μ l to 500 μ l 500 μ l to 1,000 μ l 1 ml to 2 ml 2 ml to 5 ml 5 ml to 10 ml	0.10 μ l 0.12 μ l 0.15 μ l 0.25 μ l 0.57 μ l 1.20 μ l	Calibrated using Analytical Balance and Distilled Water based on ISO 8655-6:2002 and ISO 8655-2:2002

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 22 of 40

SCOPE OF CALIBRATION: VOLUMETRIC

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks		
Viscosity Flow Cup i. Ford Cup	Cup no. 1 (10 cSt to 35 cSt)	0.07 cSt	Calibrated using Standard Solution and Stop Watch based on ASTM D 1200:2010(2014)		
	Cup no. 2 (25 cSt to 120 cSt)	0.2 cSt			
	Cup no. 3 (49 cSt to 220 cSt)	0.5 cSt			
	Cup no. 4 (70 cSt to 370 cSt)	0.6 cSt			
	Cup no. 5 (200 cSt to 1,200 cSt)	2 cSt			
	ii. Zahn Cup	Cup no.1 (5 cSt to 60 cSt)		0.1 cSt	Calibrated using Standard Solution and Stop Watch based on ASTM D 4212:2016
		Cup no. 2 (20 cSt to 250 cSt)		0.5 cSt	
		Cup no. 3 (100 cSt to 800 cSt)		1.5 cSt	
		Cup no. 4 (200 cSt to 1,200 cSt)		2 cSt	
		Cup no. 5 (400 cSt to 1,800 cSt)		3 cSt	
Hydrometer	0.600 g/ml to 1.500 g/ml	0.7 mg/ml	Compare using Hydrometer based on BS 718:1991		

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation**Signatories:**

1. **Seah Leong Ho**
2. **Kayalvili a/p Munusamy**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 23 of 40

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Tachometer (Non- Contact)	0 rpm to 1,000 rpm	1.1 rpm	Calibrated using Tachometer Calibrator and Tachometer based on ASTM F2046:2006
	1,000 rpm to 10,000 rpm	2.3 rpm	
	10,000 rpm to 20,000 rpm	4.2 rpm	
Tachometer (Contact)	0 rpm to 1,000 rpm	1.1 rpm	
	1,000 rpm to 10,000 rpm	2.3 rpm	

Signatory:

- Seah Leong Ho**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 24 of 40

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Instruments (DC & Low Frequency)			
1. DC Voltage	0 to 220 mV	8.5 μ V/V + 0.46 μ V	Fluke 5720A
	220 mV to 2.2 V	5.6 μ V/V + 0.91 μ V	
	2.2 V to 11 V	3.7 μ V/V + 5.9 μ V	
	11 V to 22 V	3.9 μ V/V + 6.3 μ V	
	22 V to 220 V	5.6 μ V/V + 59 μ V	
	220 V to 1000 V	7.2 μ V/V + 0.64 mV	
	1 kV to 10 kV	3.0 mV/V	TOS 5101, Fluke 45 & JRL KV-25
2. AC Voltage	0 to 1100 V	See Matrix A	Fluke 5720A & Wavetek 9100
3. Wideband AC Voltage	<u>1 kHz</u>		Fluke 5720A
	1.1 mV (-46 dBm)	8.0 mV/V	
	3.0 mV (-37 dBm)	7.0 mV/V	
	11 mV (-26 dBm)	7.0 mV/V	
	33 mV (-17 dBm)	6.0 mV/V	
	110 mV (-6.2 dBm)	6.0 mV/V	
	330 mV (-3.4 dBm)	5.0 mV/V	
	1.1 V (14 dBm)	5.0 mV/V	
	3.5 V (24 dBm)	4.0 mV/V	
4. DC Current	-300 μ A to -220 μ A	0.14 mA/A	Wavetek 9100
	-220 μ A to 220 μ A	70 μ A/A	Fluke 5720A
	-2.2 mA to 2.2 mA	39 μ A/A	
	-22 mA to 22 mA	37 μ A/A	
	-220 mA to 220 mA	49 μ A/A	
	-2.2 A to 2.2 A	86 μ A/A	
	-20 A to -10 A	0.55 mA/A	Wavetek 9100
	-10 A to -3 A	0.55 mA/A	
	3 A to 10 A	0.55 mA/A	
	10 A to 20 A	0.55 mA/A	
20 A to 30 A	2.0 mA/A	Yokogawa 2552 & Yokogawa 2561	
5. DC Current Clamp	10 A to 16.5 A	2.8 mA/A + 3.6 mA	Fluke 5522A & 50-Turn Coil
	16.5 A to 150 A	2.9 mA/A + 19 mA	
	150 A to 1000 A	3.2 mA/A + 17 mA	

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 25 of 40

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Instruments (DC & Low Frequency)			
6. AC Current	0 to 50 A	See Matrix B	Fluke 5720A, Wavetek 9100 & Yokogawa 2558
7. AC Current Clamp	<u>45 Hz to 65 Hz</u> 10 A to 16.5 A 16.5 A to 150 A 150 A to 1000 A	3.3 mA/A + 5.0 mA 3.5 mA/A + 32 mA 3.5 mA/A + 0.12 A	Fluke 5522A & 50-Turn Coil
	<u>65 Hz to 440 Hz</u> 10 A to 16.5 A 16.5 A to 150 A	9.3 mA/A + 5.2 mA 9.8 mA/A + 25 mA	
8. Resistance			
a) Fixed Resistance	0 Ω 1 m Ω 10 m Ω	10 $\mu\Omega/\Omega$ 0.20 m Ω/Ω 0.10 m Ω/Ω	YEW 2792
	1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω	95 $\mu\Omega/\Omega$ 23 $\mu\Omega/\Omega$ 10 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 11 $\mu\Omega/\Omega$ 20 $\mu\Omega/\Omega$ 40 $\mu\Omega/\Omega$ 0.10 m Ω/Ω	Fluke 5720A
	1.9 Ω 19 Ω 190 Ω 1.9 k Ω 19 k Ω 190 k Ω 1.9 M Ω 19 M Ω	95 $\mu\Omega/\Omega$ 23 $\mu\Omega/\Omega$ 10 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 8.5 $\mu\Omega/\Omega$ 11 $\mu\Omega/\Omega$ 21 $\mu\Omega/\Omega$ 50 $\mu\Omega/\Omega$	
b) Variable Resistance	0 to 40 Ω 40 Ω to 400 Ω 400 Ω to 4 k Ω 4 k Ω to 40 k Ω 40 k Ω to 400 k Ω 400 k Ω to 4 M Ω 4 M Ω to 40 M Ω 40 M Ω to 400 M Ω	0.25 m Ω/Ω 0.25 m Ω/Ω 0.15 m Ω/Ω 0.20 m Ω/Ω 0.20 m Ω/Ω 0.50 m Ω/Ω 1.5 m Ω/Ω 2.6 m Ω/Ω	Wavetek 9100

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 26 of 40

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Instruments (DC & Low Frequency)			
8. Resistance			
c) Earth Resistance	0.1 Ω to 1 k Ω 1 k Ω to 100 k Ω	0.10 m Ω / Ω 1.0 m Ω / Ω	Yokogawa 2793
d) LCR Resistance	1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω	0.10 m Ω / Ω 0.10 m Ω / Ω 0.10 m Ω / Ω 0.50 m Ω / Ω 0.50 m Ω / Ω 0.50 m Ω / Ω	Yokogawa 2793
e) Insulation Resistance	100 k Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω 1 G Ω to 10 G Ω 10 G Ω to 100 G Ω 100 G Ω to 1 T Ω	2.0 m Ω / Ω 2.0 m Ω / Ω 1.0 m Ω / Ω 2.0 m Ω / Ω 5.0 m Ω / Ω 5.0 m Ω / Ω	Yokogawa 2793 & IET HRRS-5kV
f) High Current Resistance			
at 200 A	50 $\mu\Omega$ 100 $\mu\Omega$ 150 $\mu\Omega$ 200 $\mu\Omega$	0.79 $\mu\Omega$ 0.45 $\mu\Omega$ 0.20 $\mu\Omega$ 1.1 $\mu\Omega$	Time Elect. 5070
at 100 A	0.5 m Ω 1.0 m Ω 1.5 m Ω 2.0 m Ω	1.4 $\mu\Omega$ 1.0 $\mu\Omega$ 1.7 $\mu\Omega$ 2.4 $\mu\Omega$	
at 30 A	5 m Ω 10 m Ω 15 m Ω 20 m Ω	11 $\mu\Omega$ 8.6 $\mu\Omega$ 11 $\mu\Omega$ 8.2 $\mu\Omega$	
at 10 A	50 m Ω 100 m Ω 150 m Ω 200 m Ω	79 $\mu\Omega$ 80 $\mu\Omega$ 81 $\mu\Omega$ 83 $\mu\Omega$	

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 27 of 40

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks	
Measuring Instruments (DC & Low Frequency)				
9. Capacitance a) Variable Capacitance	0.5 nF to 4 nF	3.4 mF/F	Wavetek 9100	
	4 nF to 40 nF	3.0 mF/F		
	40 nF to 400 nF	3.0 mF/F		
	400 nF to 4 μ F	4.0 mF/F		
	4 μ F to 40 μ F	5.0 mF/F		
	40 μ F to 400 μ F	5.0 mF/F		
	400 μ F to 4 mF	5.0 mF/F		
	4 mF to 40 mF	10 mF/F		
	b) LCR Capacitance 1 kHz	1 pF to 10 pF	0.50 mF/F	IET 1413, SC-1000, IET SC-100 & SC-10
		10 pF to 100 pF	0.50 mF/F	
		100 pF to 1 nF	0.50 mF/F	
		1 nF to 10 nF	0.50 mF/F	
		10 nF to 100 nF	0.50 mF/F	
		100 nF to 1 μ F	0.50 mF/F	
1 μ F to 10 μ F		0.20 mF/F		
10 μ F to 100 μ F		0.40 mF/F		
10. Inductance 1 kHz	100 μ H to 1 mH	20 mH/H	IET 1491-G	
	1 mH to 10 mH	20 mH/H		
	10 mH to 100 mH	20 mH/H		
	100 mH to 1 H	8.0 mH/H		
	1 H to 10 H	8.0 mH/H		
Measuring Instruments (Time & Frequency)				
11. Time	10 s to 60 s	50 ms	Pendulum 6689, HP 8662 & HP 53132A	
	60 s to 300 s	60 ms		
	300 s to 600 s	60 ms		
	600 s to 900 s	60 ms		
	900 s to 1800 s	60 ms		
	0.5 h to 1 h	90 ms		
	1 h to 3 h	0.10 s		
12. RF Power	3 μ W to 100 mW (-25 dBm to 20 dBm)	3.0 mW/W	HP 11683A, HP 432A, HP 478A & HP 34401A	
13. FM Modulation Wow & Flutter	0 to 3 kHz	3.0 %	Minato 3101 & HP 8903E	

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 28 of 40

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Instruments (Time & Frequency)			
14. Oscilloscope i) Vertical Deflection a) DC Voltage Load Impedance 1 M Ω	\pm (888 μ V to 222.4 V)	0.25 mV/V	Wavetek 9500
b) Square Voltage Load Impedance 1 M Ω	35.52 μ V _{pp} to 999.9 μ V _{pp} 1 mV _{pp} to 21 mV _{pp} 21.001mV _{pp} to 556mV _{pp} 556.01 mV _{pp} to 210V _{pp}	10 mV _{pp} /V _{pp} 1.0 mV _{pp} /V _{pp} 1.0 mV _{pp} /V _{pp} 0.50 mV _{pp} /V _{pp}	
50 Ω	35.52 μ V _{pp} to 999.9 μ V _{pp} 1 mV _{pp} to 21 mV _{pp} 21.001mV _{pp} to 556mV _{pp} 556.01 mV _{pp} to 5.56V _{pp}	10 mV _{pp} /V _{pp} 1.0 mV _{pp} /V _{pp} 1.0 mV _{pp} /V _{pp} 0.50 mV _{pp} /V _{pp}	
ii) Horizontal Deflection a) Low Edge Load Impedance 50 Ω / 1 M Ω	4.44 mV _{pp} to 3.31 V _{pp} (Rise&Fall Time : 500ps)	30 mV _{pp} /V _{pp}	
b) High Edge Load Impedance 50 Ω	888 mV _{pp} to 5.56 V _{pp} (Rise&Fall Time : 100ns)	30 mV _{pp} /V _{pp}	
1 M Ω	888 mV _{pp} to 100 V _{pp} (Rise&Fall Time : 150ns)	30 mV _{pp} /V _{pp}	
c) Fast Edge Load Impedance 50 Ω	100 V _{pp} to 210 V _{pp} (Rise&Fall Time : 200ns)	30 mV _{pp} /V _{pp}	
	4.44 mV _{pp} to 3.1 V _{pp} (Rise&Fall Time : 150ps)	30 mV _{pp} /V _{pp}	
d) Time Markers Sine (50 Ω /1 M Ω)	450.5 ps to 909.09 ps 909.1 ps to 9.009 ns	0.25 μ s/s 0.25 μ s/s	
Square (50 Ω /1 M Ω)	9.0091 ns to 55 s	0.25 μ s/s	

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 29 of 40

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Instruments (Time & Frequency)			
14. Oscilloscope ii) Horizontal Deflection a) Sine Voltage Load Impedance 50 Ω / 1 M Ω	<u>4.44 mV_{pp} to 5.56 V_{pp}</u> 50 Hz to 10 MHz 10 MHz to 100 MHz 100 MHz to 550 MHz	15 mV _{pp} /V _{pp} 15 mV _{pp} /V _{pp} 30 mV _{pp} /V _{pp}	Wavetek 9500
	<u>4.44 mV_{pp} to 3.336 V_{pp}</u> 550 MHz to 1.1 GHz	40 mV _{pp} /V _{pp}	
iii) Auxiliary a) DC Current	\pm (88.8 μ A to 111.2 mA)	2.5 mA/A	Wavetek 9100
b) Square Current	88.8 μ A _{pp} to 111.2 mA _{pp}	2.5 mA _{pp} /A _{pp}	
c) Resistance	50 k Ω to 12 M Ω 800 k Ω to 1.2 M Ω	5.0 m Ω / Ω 1.0 m Ω / Ω	
d) Capacitance	1 pF to 35 pF 35 pF to 95 pF	20 mF/F 30 mF/F	
15. Frequency	0.5 Hz to 10 MHz	25 μ Hz /Hz	

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 30 of 40

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Generating Instruments (DC & Low Frequency)			
1. DC Voltage	0 to 200 mV	5.6 μ V/V + 0.12 μ V	Fluke 8508A
	200 mV to 2 V	3.9 μ V/V + 0.58 μ V	
	2 V to 20 V	3.9 μ V/V + 5.6 μ V	
	20 V to 200 V	6.3 μ V/V + 50 μ V	
	200 V to 1000 V	6.1 μ V/V + 0.68 mV	
	0.5 kV to 10 kV	5.0 mV/V	Kikusui 149-10A
	10 kV to 25 kV	2.4 mV/V	JRL KV-25 & Fluke 45
	25 kV to 100 kV	1.5 mV/V	JRL HV-100 & Fluke 45
2. AC Voltage	0 to 50 kV	See Matrix C	Fluke 5790A, Kikusui 149-10A, JRL HV-100 & HP 34401A
3. DC Current	0 to 200 μ A	13 μ A/A + 0.64 nA	Fluke 8508A
	200 μ A to 2 mA	13 μ A/A + 6.4 nA	
	2 mA to 20 mA	15 μ A/A + 63 nA	
	20 mA to 200 mA	54 μ A/A + 0.98 μ A	
	200 mA to 2A	0.21 mA/A + 18 μ A	
	-20 A to 20 A	0.20 mA/A	Valhalla 2575A & HP 34401A
	-100 A to 100 A	0.50 mA/A	
4. AC Current	0 to 100 A	See Matrix D	Wavetek 1281, Valhalla 2575A & HP 34401A
5. Resistance	0 to 2 Ω	7.3 $\mu\Omega/\Omega$ + 55 $\mu\Omega$	Fluke 8508A
	2 Ω to 20 Ω	9.4 $\mu\Omega/\Omega$ + 45 $\mu\Omega$	
	20 Ω to 200 Ω	9.0 $\mu\Omega/\Omega$ + 62 $\mu\Omega$	
	200 Ω to 2 k Ω	9.1 $\mu\Omega/\Omega$ + 0.58 m Ω	
	2 k Ω to 20 k Ω	9.1 $\mu\Omega/\Omega$ + 5.8 m Ω	
	20 k Ω to 200 k Ω	9.1 $\mu\Omega/\Omega$ + 58 m Ω	
	200 k Ω to 2 M Ω	10 $\mu\Omega/\Omega$ + 1.1 Ω	
	2 M Ω to 20 M Ω	23 $\mu\Omega/\Omega$ + 0.11 k Ω	
	20 M Ω to 100 M Ω	0.14 m Ω/Ω + 11 k Ω	

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 31 of 40

SCOPE OF CALIBRATION: ELECTRICAL

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Generating Instruments (DC & Low Frequency)			
6. Capacitance 1 kHz	1 nF to 10 nF 10 nF to 100 nF 100 nF to 1 μ F	0.50 mF/F 0.50 mF/F 0.50 mF/F	GW LCR-819
7. Inductance 1 kHz	100 μ H to 1 mH 1 mH to 10 mH 10 mH to 100 mH	3.0 mH/H 3.0 mH/H 3.0 mH/H	GW LCR-819
Generating Instruments (Time & Frequency)			
8. Frequency	0.1 Hz to 3 GHz	0.2 μ Hz/Hz	HP 53132A
9. Time	0.33 ns to 10 s	0.20 μ s/s	HP 53132A

Signatories:

1. **Seah Leong Ho**
2. **Chin Inn Nkot**
3. **Shah Zulkifli Nor Bin Arshad**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 32 of 40

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Instruments (DC & Low Frequency)			
1. DC Voltage	\pm (0 to 300 mV) \pm (300 mV to 3 V) \pm (3 V to 30 V) \pm (30 V to 300 V) \pm (300 V to 1000 V)	0.10 mV/V 0.10 mV/V 0.10 mV/V 0.10 mV/V 0.10 mV/V	Wavetek 9100
2. AC Voltage	0 to 1000 V	See Matrix E	Wavetek 9100
3. DC Current	\pm (0 to 300 μ A) \pm (300 μ A to 3 mA) \pm (3 mA to 30 mA) \pm (30 mA to 300 mA) \pm (300 mA to 3 A) \pm (3 A to 10 A) \pm (10 A to 20 A)	0.20 mA/A 0.20 mA/A 0.20 mA/A 0.20 mA/A 0.60 mA/A 0.60 mA/A 0.60 mA/A	Wavetek 9100
4. DC Current Clamp a) X 10 Coil	3.2 A to 32 A 32 A to 105 A 105 A to 200 A	3.0 mA/A 3.0 mA/A 3.0 mA/A	Wavetek 9100
b) X 50 Coil	16 A to 160 A 160 A to 525 A 525 A to 1000 A	3.0 mA/A 3.0 mA/A 3.0 mA/A	
5. AC Current	0 to 20 A	See Matrix F	Wavetek 9100
6. AC Current Clamp a) X 10 Coil	<u>3 A to 30 A</u> 10 Hz to 100 Hz 100 Hz to 440 Hz	5.0 mA/A 10 mA/A	Wavetek 9100
b) X 50 Coil	<u>30 A to 200 A</u> 10 Hz to 100 Hz 100 Hz to 440 Hz	5.0 mA/A 10 mA/A	
	<u>16 A to 160 A</u> 10 Hz to 100 Hz	5.0 mA/A	
	<u>160 A to 1000 A</u> 10 Hz to 100 Hz	5.0 mA/A	

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 33 of 40

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks	
Measuring Instruments (DC & Low Frequency)				
7. Resistance a) Variable Resistance	0.1 Ω to 40 Ω	1.0 m Ω / Ω	Wavetek 9100	
	40 Ω to 400 Ω	0.40 m Ω / Ω		
	400 Ω to 4 k Ω	0.40 m Ω / Ω		
	4 k Ω to 40 k Ω	0.30 m Ω / Ω		
	40 k Ω to 400 k Ω	0.30 m Ω / Ω		
	400 k Ω to 4 M Ω	0.50 m Ω / Ω		
	4 M Ω to 40 M Ω	2.0 m Ω / Ω		
	40 M Ω to 400 M Ω	3.0 m Ω / Ω		
	b) LCR Resistance	1 Ω to 10 Ω	1.0 m Ω / Ω	Yokogawa 2793
		10 Ω to 100 Ω	1.0 m Ω / Ω	
		100 Ω to 1 k Ω	1.0 m Ω / Ω	
		1 k Ω to 10 k Ω	1.0 m Ω / Ω	
		10 k Ω to 100 k Ω	1.0 m Ω / Ω	
c) Insulation Resistance	100 k Ω to 1 M Ω	1.0 m Ω / Ω	Yokogawa 2793 & IET HRRS-5kV	
	10 M Ω to 100 M Ω	5.0 m Ω / Ω		
	100 M Ω to 1 G Ω	5.0 m Ω / Ω		
	1 G Ω to 10 G Ω	5.0 m Ω / Ω		
8. Capacitance a) Variable Capacitance	10 G Ω to 100 G Ω	10 m Ω / Ω	Yokogawa 2793 & IET HRRS-5kV	
	100 G Ω to 1 T Ω	10 m Ω / Ω		
	a) Variable Capacitance	0.5 nF to 4 nF	3.0 mF/F	Wavetek 9100
		4 nF to 40 nF	3.0 mF/F	
		40 nF to 400 nF	3.0 mF/F	
		400 nF to 4 μ F	4.0 mF/F	
		4 μ F to 40 μ F	10 mF/F	
		40 μ F to 400 μ F	10 mF/F	
		400 μ F to 4 mF	10 mF/F	
		4 mF to 40 mF	10 mF/F	
	b) LCR Capacitance 1 kHz	1 pF to 10 pF	1.0 mF/F	IET 1413, IET 1412-BC, IET SC-100, IET SC-100 & IET SC-1000
		10 pF to 100 pF	1.0 mF/F	
		100 pF to 1 nF	1.0 mF/F	
1 nF to 10 nF		1.0 mF/F		
10 nF to 100 nF		1.0 mF/F		
100 nF to 1 μ F		1.0 mF/F		
1 μ F to 10 μ F		1.0 mF/F		
10 μ F to 100 μ F		1.0 mF/F		
100 μ F to 1 mF		10 mF/F		

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 34 of 40

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Measuring Instruments (DC & Low Frequency)			
9. Inductance 1 kHz	100 μ H to 1 mH 1 mH to 10 mH 10 mH to 100 mH 100 mH to 1 H 1 H to 10 H	30 mH/H 30 mH/H 30 mH/H 10 mH/H 10 mH/H	IET 1491-D/G
10. Frequency	0.5 Hz to 10 MHz	25 μ Hz/Hz	Wavetek 9100
Measuring Instruments (Time & Frequency)			
11. Oscilloscope i) Vertical Deflection a) Square Wave 50 Ω 1 M Ω b) DC Level 50 Ω 1 M Ω ii) Horizontal Deflection a. Time Markers 50 Ω b. Edge Response (Rise/Fall Time) 50 Ω load 1 M Ω c. Bandwidth 50 Ω 1 M Ω	4.44 mV _{pp} to 3.34 V _{pp} 4.44 mV _{pp} to 133.44 V _{pp} \pm (4.44 mV to 2.78 V) \pm (4.44 mV to 133.44 V) 4 ns/div to 5.5 s/div < 1 ns < 100 ns 50 kHz to 100 MHz 100 MHz to 250 MHz 10 Hz to 49.999 kHz	3.0 mV/V _{pp} 3.0 mV/V _{pp} 3.0 mV/V 3.0 mV/V 25 μ s/s 25 μ s/s 25 μ s/s 2.0 mHz/Hz 4.0 mHz/Hz 3.0 mHz/Hz	Wavetek 9100

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 35 of 40

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Generating Instruments (DC & Low Frequency)			
1. DC Voltage	$\pm(0 \text{ to } 100 \text{ mV})$	0.10 mV/V	HP 34401A
	$\pm(100 \text{ mV to } 1 \text{ V})$	0.10 mV/V	
	$\pm(1 \text{ V to } 10 \text{ V})$	0.10 mV/V	
	$\pm(10 \text{ V to } 100 \text{ V})$	0.10 mV/V	
	$\pm(100 \text{ V to } 1000 \text{ V})$	0.10 mV/V	
	0.2 kV to 5 kV	10 mV/V	Kikusui 149-10A
	5 kV to 10 kV	10 mV/V	
2. AC Voltage	<u>0.5 kV to 10 kV</u> 50 Hz to 60 Hz	15 mV/V	Kikusui 149-10A
3. DC Current	$\pm(0 \text{ to } 10 \text{ mA})$	1.0 mA/A	HP 34401A
	$\pm(10 \text{ mA to } 100 \text{ mA})$	1.0 mA/A	
	$\pm(100 \text{ mA to } 1 \text{ A})$	2.0 mA/A	
	$\pm(1 \text{ A to } 3 \text{ A})$	2.0 mA/A	

Signatories:

1. **Seah Leong Ho**
2. **Chin Inn Nkot**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 36 of 40

SCOPE OF CALIBRATION: ELECTRICAL**PERMANENT & SITE CALIBRATION: CATEGORY I****MATRIX A TABLE** (AC Voltage Measuring Instruments)

i) Generated by using Fluke 5720A									
Voltage Range	Frequency								
	15 Hz to 50 Hz	20 Hz to 40 Hz	40 Hz to 20 kHz	50 Hz to 1 kHz	20 kHz to 50 kHz	50 kHz to 100 kHz	100 kHz to 300 kHz	300 kHz to 500 kHz	500 kHz to 1 MHz
0 mV to 2.2 mV	-	-	0.20 mV/V	-	0.50 mV/V	-	-	-	-
2.2 mV to 22 mV	-	0.28 mV/V	0.28 mV/V	-	0.40 mV/V	0.75 mV/V	1.5 mV/V	2.3 mV/V	3.6 mV/V
22 mV to 220 mV	-	0.12 mV/V	0.11 mV/V	-	0.23 mV/V	0.54 mV/V	0.99 mV/V	1.5 mV/V	2.9 mV/V
220 mV to 2.2 V	-	0.10 mV/V	50 μ V/V	-	80 μ V/V	0.12 mV/V	0.46 mV/V	1.1 mV/V	1.8 mV/V
2.2 V to 22 V	-	97 μ V/V	48 μ V/V	-	80 μ V/V	0.11 mV/V	0.30 mV/V	1.1 mV/V	1.7 mV/V
22 V to 220 V	-	90 μ V/V	52 μ V/V	-	80 μ V/V	0.15 mV/V	0.90 mV/V	-	-
220 V to 1100 V	0.30 mV/V	-	-	70 μ V/V	-	-	-	-	-

ii) Generated by using Wavetek 9100									
Voltage Range	Frequency								
	1 kHz to 3 kHz	3 kHz to 10 kHz	10 kHz to 20 kHz	-	-	-	-	-	-
220 V to 1100 V	0.80 mV/V	0.80 mV/V	1.2 mV/V	-	-	-	-	-	-

MATRIX B TABLE (AC Current Measuring Instruments)

i) Generated by using Fluke 5720A				
Current Range	Frequency			
	20 Hz to 1 kHz	40 Hz to 1 kHz	1 kHz to 5 kHz	5 kHz to 10 kHz
30 μ A to 220 μ A	-	0.16 mA/A	0.34 mA/A	1.4 mA/A
220 μ A to 2.2 mA	-	0.14 mA/A	0.25 mA/A	1.4 mA/A
2.2 mA to 22 mA	-	0.13 mA/A	0.21 mA/A	1.2 mA/A
22 mA to 2.2 A	0.28 mA/A	-	0.49 mA/A	7.1 mA/A

ii) Generated by using Wavetek 9100				
Current Range	Frequency			
	10 Hz to 3 kHz	3 kHz to 10 kHz	10 kHz to 30 kHz	-
0 μ A to 30 μ A	0.70 mA/A	1.0 mA/A	2.0 mA/A	-
3 A to 10 A	2.0 mA/A	5.0 mA/A	-	-
10 A to 20 A	2.0 mA/A	5.0 mA/A	-	-

iii) Generated by using Yokogawa 2558				
Current Range	Frequency			
	50 Hz to 60 Hz	60 Hz to 400 Hz	-	-
20 A to 50 A	0.15 mA/A	0.15 mA/A	-	-

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 37 of 40

SCOPE OF CALIBRATION: ELECTRICAL**PERMANENT & SITE CALIBRATION: CATEGORY I****MATRIX C TABLE** (AC Voltage Generating Instruments)

i) Measured by using Fluke 5790A			
Voltage Range	Frequency		
	45 Hz to 20 Hz	20 kHz to 50 kHz	50 kHz to 100 kHz
7 mV to 22 mV	0.12 mV/V + 1.6 μ V	0.24 mV/V + 2.3 μ V	0.35 mV/V + 2.9 μ V
22 mV to 70 mV	78 μ V/V + 1.8 μ V	0.15 mV/V + 2.3 μ V	0.29 mV/V + 2.9 μ V
70 mV to 220 mV	49 μ V/V + 1.7 μ V	83 μ V/V + 2.3 μ V	0.18 mV/V + 2.8 μ V
220 mV to 700 mV	44 μ V/V + 1.7 μ V	64 μ V/V + 2.2 μ V	96 μ V/V + 2.8 μ V
700 mV to 2.2 V	32 μ V/V	59 μ V/V	86 μ V/V
2.2 V to 7 V	34 μ V/V	61 μ V/V	0.1 mV/V
7 V to 22 V	35 μ V/V	60 μ V/V	96 μ V/V
22 V to 70 V	45 μ V/V	72 μ V/V	0.12 mV/V
70 V to 220 V	43 μ V/V	87 μ V/V	0.12 mV/V
220 V to 700 V	54 μ V/V	-	-
700 V to 1000 V	49 μ V/V	-	-
ii) Measured by using Kikusui 149-10A			
Voltage Range	Frequency		
	50 Hz to 60 Hz	-	-
0.5 kV to 10 kV	10 mv/V	-	-
iii) Measured by using JRL HV-100 & HP 34401A			
Voltage Range	Frequency		
	50 Hz to 60 Hz	-	-
10 kV to 25 kV	0.40 V/V	-	-
25 kV to 50 kV	0.80 V/V	-	-

MATRIX D TABLE (AC Current Generating Instruments)

i) Measured by using Fluke 8508A			
Current Range	Frequency		
	40 Hz to 2 kHz	45 Hz to 10 kHz	2 kHz to 10 kHz
0 to 200 μ A	-	0.56 mA/A + 23 nA	-
200 μ A to 2 mA	-	0.34 mA/A + 0.23 μ A	-
2 mA to 20 mA	-	0.34 mA/A + 2.3 μ A	-
20 mA to 200 mA	-	0.33 mA/A + 23 μ A	-
200 mA to 2 A	0.70 mA/A + 0.23 mA	-	0.83 mA/A + 0.23 mA
ii) Measured by using Valhalla 2575A & HP 34401A			
Current Range	Frequency		
	50 Hz to 1 kHz	1 kHz to 10 kHz	-
2 A to 20 A	1.0 mA/A	5.0 mA/A	-
20 A to 100 A	1.0 mA/A	5.0 mA/A	-

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 38 of 40

SCOPE OF CALIBRATION: ELECTRICAL**PERMANENT & SITE CALIBRATION: CATEGORY I****MATRIX E TABLE** (AC Voltage Measuring Instruments)

Generated by using Wavetek 9100						
Voltage Range	Frequency					
	10 Hz to 3 kHz	3 kHz to 10 kHz	10 kHz to 30 kHz	30 kHz to 50 kHz	50 kHz to 100 kHz	-
0 mV to 10 mV	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	5.0 mV/V	-
10 mV to 30 mV	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	5.0 mV/V	-
30 mV to 300 mV	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	5.0 mV/V	-
300 mV to 1.5 V	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	2.0 mV/V	-
1.5 V to 3 V	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	2.0 mV/V	-
3 V to 30 V	1.0 mV/V	1.0 mV/V	1.0 mV/V	2.0 mV/V	4.0 mV/V	-
30 V to 100 V	1.0 mV/V	1.0 mV/V	1.0 mV/V	2.0 mV/V	4.0 mV/V	-
Voltage Range	Frequency					
	40 Hz to 100 Hz	100 Hz to 1 kHz	1 kHz to 3 kHz	3 kHz to 10 kHz	10 kHz to 20 kHz	20 kHz to 30 kHz
100 V to 300 V	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	2.0 mV/V
300 V to 750 V	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	2.0 mV/V
750 V to 1000 V	1.0 mV/V	1.0 mV/V	1.0 mV/V	1.0 mV/V	2.0 mV/V	-

MATRIX F TABLE (AC Current Measuring Instruments)

Generated by using Wavetek 9100				
Current Range	Frequency			
	10 Hz to 3 kHz	3 kHz to 10 kHz	10 kHz to 20 kHz	20 kHz to 30 kHz
0 to 30 μ A	1.0 mA/A	2.0 mA/A	3.0 mA/A	3.0 mA/A
30 μ A to 300 μ A	1.0 mA/A	2.0 mA/A	3.0 mA/A	3.0 mA/A
300 μ A to 3 mA	1.0 mA/A	1.0 mA/A	2.0 mA/A	3.0 mA/A
3 mA to 30 mA	1.0 mA/A	1.0 mA/A	2.0 mA/A	3.0 mA/A
30 mA to 300 mA	1.0 mA/A	1.0 mA/A	2.0 mA/A	3.0 mA/A
300 mA to 3 A	2.0 mA/A	3.0 mA/A	2.0 mA/A	3.0 mA/A
3 A to 10 A	3.0 mA/A	10 mA/A	-	-
10 A to 20 A	3.0 mA/A	10 mA/A	-	-

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 39 of 40

SCOPE OF CALIBRATION: OPTICAL AND PHOTOMETRIC MEASUREMENTS

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Spectrophotometer (UV- Visible Range)			
a) Wavelength	240 nm to 645 nm	0.2 nm	Calibrated using Holmium perchlorate as standard based on ASTM E275:2008(2013) and ASTM E925:2009(2014)
b) Absorbance (230 nm to 640 nm)	0.2 to 1.0	0.006	Calibrated using Potassium Dichromate and Neutral Density Filter as standard based on ASTM E275:2008(2013) and ASTM E925:2009(2014)
Transmittance (Normal incidence)	240 nm to 780 nm	0.94 %	Calibrated using Agilent Cary 7000 Spectrophotometer based on ASTM D1746:2015 and ISO 9050:2003

Scan QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation**Signatories:**

1. **Seah Leong Ho**
2. **Teo Kok Siong**

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 40 of 40

SCOPE OF CALIBRATION: OPTICAL AND PHOTOMETRIC MEASUREMENTS**SITE: CATEGORY I**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Spectrophotometer (UV-Visible Range)			
a) Wavelength	240 nm to 645 nm	0.2 nm	Calibrated using Holmium perchlorate as standard based on ASTM E275:2008(2013) and ASTM E925:2009(2014)
b) Absorbance (230 nm to 640 nm)	0.2 to 1.0	0.006	Calibrated using Potassium Dichromate and Neutral Density Filter as standard based on ASTM E275:2008(2013) and ASTM E925:2009(2014)

Signatories:

1. **Seah Leong Ho**
2. **Teo Kok Siong**