

Schedule

Issue date: 3 January 2017
Valid until: 23 November 2017



NO: SAMM 082

(Issue 3, 3 January 2017 replacement of SAMM 082 dated 12 January 2016)

Page: 12 of 46

FIELD OF CALIBRATION: MASS

SCOPE OF ACCREDITATION:

The valid scope of accreditation is in www.ism.gov.my/cab-directories.

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Standard Weight	1 mg	0.003 mg	Calibrated using Standard Weight Sets and Mass Comparator based on OIML R111-2:2004
	2 mg	0.003 mg	
	5 mg	0.003 mg	
	10 mg	0.003 mg	
	20 mg	0.003 mg	
	50 mg	0.003 mg	
	100 mg	0.003 mg	
	200 mg	0.003 mg	
	500 mg	0.004 mg	
	1 g	0.004 mg	
	2 g	0.007 mg	
	5 g	0.01 mg	
	10 g	0.02 mg	
	20 g	0.03 mg	
	50 g	0.06 mg	
	100 g	0.13 mg	
	200 g	0.26 mg	
	500 g	0.67 mg	
	1 kg	1.3 mg	
	2 kg	2.6 mg	
5 kg	6.7 mg		
10 kg	0.26g		
20 kg	0.52 g		
25 kg	1.46 g		

Signatory:

1. **Seah Leong Ho**

Schedule

Issue date: 3 January 2017
Valid until: 23 November 2017



NO: SAMM 082

(Issue 3, 3 January 2017 replacement of SAMM 082 dated 12 January 2016)

FIELD OF CALIBRATION: MASS

SITE CALIBRATION: CATE OR I

SCOPE OF ACCREDITATION:

The valid scope of accreditation is in www.ism.gov.my/cab-directories.

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(±)*	Remarks
Analytical Balance and Weighing Scale	0 g to 5 g	0.02 mg	Calibrated using Standard Weight Sets based on ASTM E898:2005
	0 g to 20 g	0.02 mg	
	0 g to 50 g	0.15 mg	
	0 g to 100 g	0.11 mg	
	0 g to 200 g	0.20 mg	
	0 g to 500 g	0.50 mg	
	0 kg to 1 kg	1.0 mg	
	0 kg to 2 kg	2.0 mg	
	0 kg to 3 kg	2.2 mg	
	0 kg to 5 kg	5.0 mg	
	0 kg to 10 kg	47 mg	
	0 kg to 20 kg	96 mg	
	0 kg to 30 kg	1.07 g	
	0 kg to 50 kg	2.0 g	
	0 kg to 100 kg	5.0 g	
	0 kg to 200 kg	9.0 g	
	0 kg to 500 kg	20 g	
	0 kg to 750 kg	45 g	
	0 kg to 1000 kg	47 g	
	0 kg to 1500 kg	54 g	
0 kg to 2000 kg	94 g		
Standard Weight	1 kg	14 mg	Calibrated using Standard Weight Sets and Mass Comparator based on OIML R111-2:2004
	2 kg	14 mg	
	5 kg	16 mg	
	10 kg	0.14 g	
	20 kg	0.15 g	
	25 kg	0.15 g	

Signatory:

1. **Seah Leong Ho**