

NO: SAMM 082(Issue 2, 18 May 2018 replacement
of SAMM 082 dated 8 December 2017)

Page: 15 of 40

SCOPE OF CALIBRATION: FLOW

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty(\pm)*	Remarks
Anemometer	0.15 m/s to 5 m/s	0.12 m/s	Calibrated using Anemometer Calibrator
	5 m/s to 20 m/s	0.24 m/s	
	20 m/s to 45 m/s	0.54 m/s	

Signatory:

1. **Seah Leong Ho**