

## Schedule

Issue date: 3 January 2017  
Valid until: 23 November 2017



### NO: SMM 082

(Issue 3, 3 January 2017 replacement  
of SMM 082 dated 12 January 2016)

Page: 15 of 46

**FIELD OF CALIBRATION: FLOW**

**SCOPE OF ACCREDITATION:**

Instrument Calibrated/ Measurement Parameter	Range	Calibration and Measurement Capability Expressed as an Uncertainty( $\pm$ )*	Remarks
Anemometer	0.15 m/s to 5 m/s	0.12 m/s	Calibrated using Anemometer Calibrator
	5 m/s to 20 m/s	0.24 m/s	
	20 m/s to 45 m/s	0.54 m/s	

The valid scope of accreditation is in [www.ism.gov.my/cab-directories](http://www.ism.gov.my/cab-directories).

**Signatory:**

1. **Seah Leong Ho**